



Division of Criminal Justice Services

CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS MAINTAINED IN THE REGULAR COURSE OF BUSINESS

I, Joshua Vinehout, Supervisor of Public Safety Programs, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NAT'L PATENT ANALYTICAL SYS, DATAMASTER DMT breath test instrument, serial number 117106, performed on May 05, 2021 by Office of Public Safety employee Martin Schilling, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/ maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Joshua Vinehout".

Digitally signed under
ESRA by Joshua Vinehout
on 5/13/2021 2:45 PM

Joshua Vinehout
Supervisor of Public Safety Programs
Office of Public Safety

An Equal Opportunity/Affirmative Action Agency

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Glens Falls City Police Department

Manufacturer: NAT'L PATENT ANALYTICAL SYS Model: DATAMASTER DMT

Date: May 05, 2021

Serial Number: 117106

Reference Solution Lot Number: 20361

Reference Solution Standard: 0.10 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.098 %	34.01 degrees centigrade
2	0.098 %	34.01 degrees centigrade
3	0.098 %	34.01 degrees centigrade
4	0.098 %	34.00 degrees centigrade

Reference Solution Lot Number: 20470

Reference Solution Standard: 0.08 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.079 %	34.00 degrees centigrade
2	0.079 %	34.01 degrees centigrade
3	0.079 %	34.01 degrees centigrade
4	0.079 %	34.01 degrees centigrade

Reference Solution Lot Number: 20290

Reference Solution Standard: 0.180 % at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.177 %	34.01 degrees centigrade
2	0.178 %	34.01 degrees centigrade
3	0.178 %	34.01 degrees centigrade
4	0.178 %	34.01 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NAT'L PATENT ANALYTICAL SYS breath test instrument model DATAMASTER DMT, serial number 117106, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under
ESRA by Martin Schilling
on 5/5/2021 11:01 AM

Martin Schilling
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 45926

Reviewed by:



Digitally signed under
ESRA by Frank Sgarlata
on 5/10/2021 8:42 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Glens Falls City Police Department
Manufacturer: NAT'L PATENT ANALYTICAL SYS
Model: DATAMASTER DMT
Serial Number: 117106
Date: May 05, 2021

1. General Appearance and condition of this device -----: OK

2. Check controls and indicators -----: OK

3. Check breath flow system -----: OK

4. Check printer operation -----: OK

5. System cycles through all modes -----: OK

6. Record any repairs made to this instrument and list services provided: INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES PRINT AND REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE, RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION



Performed by:

Digitally signed under
ESRA by Martin Schilling
on 5/5/2021 11:01 AM

Martin Schilling
Highway Safety Equipment Technician
Breath Analysis Operator Permit # 45926

Reviewed by:

Digitally signed under
ESRA by Frank Sgarlata
on 5/10/2021 8:42 AM

Frank Sgarlata
Breath Analysis Technical Supervisor

TECHNICIAN TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106





Date: 05/05/2021

OPERATOR NAME:
 MARTIN J SCHILLING
 Permit #: 45926
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 20361

Nominal = 0.100 g/210L

BLANK TEST	0.000	10:10
INTERNAL STANDARD	VERIFIED	10:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.098	10:10
BLANK TEST	0.000	10:11
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.098	10:11
BLANK TEST	0.000	10:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.098	10:13
BLANK TEST	0.000	10:14
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.098	10:14
BLANK TEST	0.000	10:15

Average = 0.0980
 Std Dev = 0.0000



TECHNICIAN TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106





Date: 05/05/2021

OPERATOR NAME:
MARTIN J SCHILLING
Permit #: 45926
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20470

Nominal = 0.080 g/210L

BLANK TEST	0.000	10:29
INTERNAL STANDARD	VERIFIED	10:29
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.079	10:30
BLANK TEST	0.000	10:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.079	10:31
BLANK TEST	0.000	10:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.079	10:32
BLANK TEST	0.000	10:33
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.079	10:33
BLANK TEST	0.000	10:34

Average = 0.0790
Std Dev = 0.0000



TECHNICIAN TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/05/2021

OPERATOR NAME:
MARTIN J SCHILLING
Permit #: 45926
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20290

Nominal = 0.180 g/210L

BLANK TEST	0.000	10:37
INTERNAL STANDARD	VERIFIED	10:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.177	10:38
BLANK TEST	0.000	10:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.178	10:39
BLANK TEST	0.000	10:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.178	10:41
BLANK TEST	0.000	10:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.178	10:42
BLANK TEST	0.000	10:43

Average = 0.1778
Std Dev = 0.0005

