



**CERTIFICATION PURSUANT TO CPLR 4518 OF RECORDS
MAINTAINED IN THE REGULAR COURSE OF BUSINESS**

I, **Steven J. Carluccio**, Highway Safety Equipment Technical Supervisor, Office of Public Safety, New York State Division of Criminal Justice Services, 80 South Swan Street, Albany, New York 12210, having been duly designated and authorized by the Commissioner of the Division of Criminal Justice Services, do hereby certify and authenticate, as provided by subdivision c of Rule 4518 of the Civil Practice Law and Rules, that the document annexed hereto is an exact copy of an electronic record of the Office of Public Safety which is in my possession, custody and control. This record, once created, is stored in an electronic format that cannot thereafter be altered or modified.

I further certify that the original electronic record of inspection/maintenance/calibration of NATL PATENT ANALYTICAL SYSTEM, DATAMASTER DMT breath test instrument, serial number 117106, performed on May 09, 2022 by Office of Public Safety employee George Coffin, of which the annexed is an exact copy, was made in the regular course of business of the New York State Division of Criminal Justice Services, Office of Public Safety, that such record was made at the time such inspection/maintenance/calibration was performed or within a reasonable time thereafter, and further that it is the regular course of the Office of Public Safety's business to make such records at the time such inspection/maintenance /calibration is performed, or within a reasonable time thereafter, and to provide such instrument's records to the agency that requested them.

A handwritten signature in black ink that reads "Steven J. Carluccio".

Digitally signed under ESRA
by Steven J. Carluccio
on 05/20/2022 06:58 AM

Steven J. Carluccio
Highway Safety Equipment Technical Supervisor
Office of Public Safety

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Glens Falls City Police Department

Manufacturer: NATL PATENT ANALYTICAL SYSTEM

Model: DATAMASTER DMT

Serial Number: 117106

Date: May 09, 2022

Reference Solution Lot Number : 21200

Reference Solution Standard : 0.10% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.099 %	34.01 degrees centigrade
2	0.098 %	34.01 degrees centigrade
3	0.099 %	34.01 degrees centigrade
4	0.099 %	34.01 degrees centigrade

Reference Solution Lot Number : 21310

Reference Solution Standard : 0.08% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.079 %	34.00 degrees centigrade
2	0.079 %	34.01 degrees centigrade
3	0.079 %	34.00 degrees centigrade
4	0.079 %	34.01 degrees centigrade

Reference Solution Lot Number : 22060

Reference Solution Standard : 0.18% at 34.0 degrees centigrade

Test No	Test Result	Simulator Temperature
1	0.179 %	34.01 degrees centigrade
2	0.179 %	34.00 degrees centigrade
3	0.179 %	34.01 degrees centigrade
4	0.178 %	34.01 degrees centigrade

I hereby certify that I have performed any necessary maintenance procedures and calibrated NATL PATENT ANALYTICAL SYSTEM breath test instrument model DATAMASTER DMT, serial number 117106, and have determined that it is accurate and reliable for the determination of ethyl alcohol in the blood by analysis of the breath. The instrument's test results are recorded above and this instrument satisfies all limits and standards established by the New York State Commissioner of Health as enumerated in Part 59, Subchapter D of Chapter II, TITLE 10 (Health) of the Official Compilation of Codes, Rules and Regulations of the State of New York, (10 NYCRR Part 59) as amended.

I further certify that I made the entries in these records of inspection/maintenance/calibration at the time that such inspection/maintenance/calibration of the above identified breath test instrument was performed, or within a reasonable time thereafter.

Performed by:



Digitally signed under ESRA
by George Coffin
on 05/09/2022 01:48 PM

Reviewed by:



Digitally signed under ESRA
by Frank Sgarlata
on 05/09/2022 02:44 PM

George Coffin
Highway Safety Equipment Technician
Breath Analysis Operator Permit #48534

Frank Sgarlata
Breath Analysis Technical Supervisor

**BREATH TEST INSTRUMENT
RECORD OF INSPECTION / MAINTENANCE / CALIBRATION**

Name of Submitting Agency: Glens Falls City Police Department
Manufacturer: NATL PATENT ANALYTICAL SYSTEM
Model: DATAMASTER DMT
Serial Number: 117106
Date: May 09, 2022

- 1. General appearance and condition of this device -----: OK

- 2. Check controls and indicators -----: OK

- 3. Check breath flow system -----: OK

- 4. Check printer operation -----: OK

- 5. System cycles through all modes -----: OK



6. Record any repairs made to this instrument and list services provided : INSPECTED INSTRUMENT, CHECKED SETUP VOLTAGES PRINT AND REMOTE OPERATION. PERFORMED CALIBRATION/ADJUSTMENT PROCEDURE, RAN SUBJECT, TECHNICIAN AND DIAGNOSTIC TESTS, VERIFIED CALIBRATION.

Performed by:

Digitally signed under ESRA
by George Coffin
on 05/09/2022 01:48 PM

Reviewed by:

Digitally signed under ESRA
by Frank Sgarlata
on 05/09/2022 02:44 PM

George Coffin
Highway Safety Equipment Technician
Breath Analysis Operator Permit #48534

Frank Sgarlata
Breath Analysis Technical Supervisor

TECHNICIAN TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106





Date: 05/09/2022

OPERATOR NAME:
 GEORGE R COFFIN
 Permit #: 48534
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 21310

Nominal = 0.080 g/210L

BLANK TEST	0.000	12:03
INTERNAL STANDARD	VERIFIED	12:03
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.079	12:03
BLANK TEST	0.000	12:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.079	12:05
BLANK TEST	0.000	12:05
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.079	12:06
BLANK TEST	0.000	12:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.079	12:07
BLANK TEST	0.000	12:08

Average = 0.0790
 Std Dev = 0.0000



TECHNICIAN TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106





Date: 05/09/2022

OPERATOR NAME:
GEORGE R COFFIN
Permit #: 48534
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 21200

Nominal = 0.100 g/210L

BLANK TEST	0.000	12:11
INTERNAL STANDARD	VERIFIED	12:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.099	12:12
BLANK TEST	0.000	12:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.098	12:13
BLANK TEST	0.000	12:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.099	12:14
BLANK TEST	0.000	12:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.099	12:16
BLANK TEST	0.000	12:16

Average = 0.0988
Std Dev = 0.0005



TECHNICIAN TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106





Date: 05/09/2022

OPERATOR NAME:
GEORGE R COFFIN
Permit #: 48534
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 22060

Nominal = 0.180 g/210L

BLANK TEST	0.000	12:18
INTERNAL STANDARD	VERIFIED	12:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.179	12:19
BLANK TEST	0.000	12:20
REFERENCE STANDARD TEMP	34.00c	
REFERENCE STANDARD	0.179	12:20
BLANK TEST	0.000	12:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.179	12:22
BLANK TEST	0.000	12:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.178	12:23
BLANK TEST	0.000	12:24

Average = 0.1788
Std Dev = 0.0005



* 5/9/22
GRC

SERVICE AUTHORIZATION FORM
Highway Safety Technology Unit
 80 South Swan Street
 Albany, NY 12210
 Ph: 518-485-7636 / 518-402-0689
 Fax: 518-457-6869

THIS FORM IS USED TO AUTHORIZE THE DIVISION OF CRIMINAL JUSTICE SERVICES TO PERFORM SERVICE ON THE INSTRUMENT IDENTIFIED. THIS AUTHORIZATION MUST ACCOMPANY EACH PIECE OF EQUIPMENT WHEN IT IS PRESENTED TO THE EQUIPMENT REPAIR CENTER FOR SERVICE. **PLEASE SIGN AND DATE FORM.**

SECTION I - AGENCY INFORMATION

Name of Agency GLENS FALLS POLICE DEPARTMENT		Today's Date 05/03/2022
Address 42 RIDGE ST		City, State, ZIP GLENS FALLS, NY 12801
Contact Person LT MATTISON		Contact Telephone 518-761-3840 X 3782
Contact Email CMATTISON@GLENSFALLSPD.COM		Best Way to Reach You During Appointment Period PHONE/EMAIL

SECTION II - INSTRUMENT INFORMATION

Instrument Manufacturer (Please Circle) NPAS DMT DMT GUTH Applied Concepts (Stalker) Decatur (Genesis) Kustom (Eagle/Falcon) MPH Alco-Sensor Other:		
Model Number DMT	Instrument Serial Number 117106	
Antenna Serial Number (if applicable) Antenna 1 -- / Antenna 2 --	Type of maintenance (check appropriate) <input checked="" type="checkbox"/> Calibration Check <input type="checkbox"/> Repair	
Brief Description of Malfunction YEARLY CERTIFICATION-SERVICE APPOINTMENT SCHEDULED FOR 05/09/2022		
List all parts, cables and/or accessories being submitted. Tuning forks must accompany all radar units (if applicable). (Please Circle)		
DMT/SIM: <input checked="" type="checkbox"/> DMT <input checked="" type="checkbox"/> SIM <input checked="" type="checkbox"/> Tubes <input checked="" type="checkbox"/> Keyboard Other:	Radar: Control Unit Antenna(s) Antenna Cable(s) Power Cord Forks Other:	
I acknowledge I am signing this Service Authorization Form as an act and deed of said agency, organization, or municipality, and that I am duly authorized to sign same for the uses and purposes mentioned herein. PLEASE SIGN AND DATE.		
Signature  5375		Date 05/03/2022

DCJS USE ONLY

Inspected instrument, checked setup multimeter, print & remote operation. Performed calibration / adjustment procedure, ran subject, technician + diagnostic tests, reified calibration.