

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 01/01/2019

OPERATOR NAME:
 SHAWN A LOVELACE
 Permit #: 30657
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	09:16
INTERNAL STANDARD	VERIFIED	09:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:17
BLANK TEST	0.00	09:18
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:18
BLANK TEST	0.00	09:19
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	09:19
BLANK TEST	0.00	09:20

