

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 01/01/2020

OPERATOR NAME:
 GERALD C WILLETTE
 Permit #: 32550
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:05
INTERNAL STANDARD	VERIFIED	02:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:06
BLANK TEST	0.00	02:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:07
BLANK TEST	0.00	02:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:08
BLANK TEST	0.00	02:09

