

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/07/2019

OPERATOR NAME:
TYLER R MELLO
Permit #: 46781
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:21
INTERNAL STANDARD	VERIFIED	02:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:21
BLANK TEST	0.00	02:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:22
BLANK TEST	0.00	02:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	02:24
BLANK TEST	0.00	02:25

