

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/14/2019

OPERATOR NAME:
PATRICK J COYNE
Permit #: 46590
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:43
INTERNAL STANDARD	VERIFIED	00:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:44
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:44
BLANK TEST	0.00	00:45
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:46
BLANK TEST	0.00	00:47

