

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 01/21/2019

OPERATOR NAME:
TYLER R MELLO
Permit #: 46781
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:22
INTERNAL STANDARD	VERIFIED	03:22
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:23
BLANK TEST	0.00	03:24
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:24
BLANK TEST	0.00	03:25
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:26
BLANK TEST	0.00	03:26

