

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 01/28/2019

OPERATOR NAME:  
ROBERT M FOO  
Permit #: 46059  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 18310

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:12
INTERNAL STANDARD	VERIFIED	00:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:12
BLANK TEST	0.00	00:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:14
BLANK TEST	0.00	00:15
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:15
BLANK TEST	0.00	00:16

