

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/04/2019

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:56
INTERNAL STANDARD	VERIFIED	00:56
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:56
BLANK TEST	0.00	00:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:58
BLANK TEST	0.00	00:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	00:59
BLANK TEST	0.00	01:00

