

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/18/2019

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:07
INTERNAL STANDARD	VERIFIED	01:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:08
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:09
BLANK TEST	0.00	01:10
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:10
BLANK TEST	0.00	01:11

