

SOLUTION CHANGE

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 03/01/2019

OPERATOR NAME:  
GERALD C WILLETTE  
Permit #: 32550  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:30
INTERNAL STANDARD	VERIFIED	03:30
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:30
BLANK TEST	0.00	03:31
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:31
BLANK TEST	0.00	03:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:33
BLANK TEST	0.00	03:34

