

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/11/2019

OPERATOR NAME:
PATRICK J COYNE
Permit #: 46590
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:35
INTERNAL STANDARD	VERIFIED	00:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:36
BLANK TEST	0.00	00:37
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:37
BLANK TEST	0.00	00:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	00:38
BLANK TEST	0.00	00:39

