

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 03/18/2019

OPERATOR NAME:  
PATRICK J COYNE  
Permit #: 46590  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 18360

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:05
INTERNAL STANDARD	VERIFIED	01:05
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:05
BLANK TEST	0.00	01:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:06
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:08
BLANK TEST	0.00	01:08

