

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 04/01/2019

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:32
INTERNAL STANDARD	VERIFIED	01:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:33
BLANK TEST	0.00	01:34
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:34
BLANK TEST	0.00	01:35
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	01:35
BLANK TEST	0.00	01:36

