

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 04/08/2019

OPERATOR NAME:
MARC J BARRETT
Permit #: 46725
Operator Agency/Dept: GLENS FALLS POLICE
DEPARTMENT
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:09
INTERNAL STANDARD	VERIFIED	03:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:09
BLANK TEST	0.00	03:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:11
BLANK TEST	0.00	03:12
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	03:12
BLANK TEST	0.00	03:13

