

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/03/2019

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:26
INTERNAL STANDARD	VERIFIED	01:26
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:27
BLANK TEST	0.00	01:28
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:28
BLANK TEST	0.00	01:29
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	01:29
BLANK TEST	0.00	01:30

