

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/06/2019

OPERATOR NAME:
SHAWN A LOVELACE
Permit #: 30657
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	15:26
INTERNAL STANDARD	VERIFIED	15:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:26
BLANK TEST	0.00	15:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:28
BLANK TEST	0.00	15:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	15:29
BLANK TEST	0.00	15:30

