

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 05/20/2019

OPERATOR NAME:  
MARC J BARRETT  
Permit #: 46725  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 19030

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:25
INTERNAL STANDARD	VERIFIED	01:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:26
BLANK TEST	0.00	01:27
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:27
BLANK TEST	0.00	01:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:28
BLANK TEST	0.00	01:29

