

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106


Date: 06/03/2019

OPERATOR NAME:
MARC J BARRETT
Permit #: 46725
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19090

Nominal = 0.10 g/210L



BLANK TEST	0.00	01:15
INTERNAL STANDARD	VERIFIED	01:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:15
BLANK TEST	0.00	01:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:16
BLANK TEST	0.00	01:17
REFERENCE STANDARD TEMP	34.02c	
BLANK TEST	0.00	01:18

b2340 G A 7000



BLANK TEST

0.00 01:19



SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/01/2019

OPERATOR NAME:
SHAWN A LOVELACE
Permit #: 30657
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	16:38
INTERNAL STANDARD	VERIFIED	16:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:38
BLANK TEST	0.00	16:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	16:40
BLANK TEST	0.00	16:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	16:41
BLANK TEST	0.00	16:42

