

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 06/17/2019

OPERATOR NAME:  
MARC J BARRETT  
Permit #: 46725  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:03
INTERNAL STANDARD	VERIFIED	02:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:03
BLANK TEST	0.00	02:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:05
BLANK TEST	0.00	02:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:06
BLANK TEST	0.00	02:07

