

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 07/15/2019

OPERATOR NAME:  
TYLER R MELLO  
Permit #: 46781  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 19090

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:38
INTERNAL STANDARD	VERIFIED	02:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42

