

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/01/2019

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	05:26
INTERNAL STANDARD	VERIFIED	05:26
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:27
BLANK TEST	0.00	05:28
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:28
BLANK TEST	0.00	05:29
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	05:29
BLANK TEST	0.00	05:30

