

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 08/26/2019

OPERATOR NAME:
 TYLER R MELLO
 Permit #: 46781
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:01
INTERNAL STANDARD	VERIFIED	02:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:02
BLANK TEST	0.00	02:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:03
BLANK TEST	0.00	02:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:04
BLANK TEST	0.00	02:05

