

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 09/09/2019

OPERATOR NAME:
PATRICK J COYNEP
Permit #: 46590
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:03
INTERNAL STANDARD	VERIFIED	00:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:04
BLANK TEST	0.00	00:05
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:05
BLANK TEST	0.00	00:06
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:06
BLANK TEST	0.00	00:07

