

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 09/13/2019

OPERATOR NAME:
 SHAWN A LOVELACE
 Permit #: 30657
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	10:00
INTERNAL STANDARD	VERIFIED	10:00
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:00
BLANK TEST	0.00	10:01
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	10:02
BLANK TEST	0.00	10:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	10:03
BLANK TEST	0.00	10:04

