

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 09/23/2019

OPERATOR NAME:  
TYLER R MELLO  
Permit #: 46781  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:21
INTERNAL STANDARD	VERIFIED	03:21
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:21
BLANK TEST	0.00	03:22
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:23
BLANK TEST	0.00	03:24
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:24
BLANK TEST	0.00	03:25

