

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 09/30/2019

OPERATOR NAME:
TYLER R MELLO
Permit #: 46781
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19130

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:16
INTERNAL STANDARD	VERIFIED	01:16
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:17
BLANK TEST	0.00	01:18
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:18
BLANK TEST	0.00	01:19
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:19
BLANK TEST	0.00	01:20

