

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 10/21/2019

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

| | | |
|-------------------------|----------|-------|
| BLANK TEST | 0.00 | 00:18 |
| INTERNAL STANDARD | VERIFIED | 00:18 |
| REFERENCE STANDARD TEMP | 34.01c | |
| REFERENCE STANDARD | 0.09 | 00:19 |
| BLANK TEST | 0.00 | 00:19 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.10 | 00:20 |
| BLANK TEST | 0.00 | 00:21 |
| REFERENCE STANDARD TEMP | 34.02c | |
| REFERENCE STANDARD | 0.09 | 00:21 |
| BLANK TEST | 0.00 | 00:22 |

