

SUPERVISOR TEST

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 10/28/2019

OPERATOR NAME:
 TYLER R MELLO
 Permit #: 46781
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:09
INTERNAL STANDARD	VERIFIED	02:09
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:09
BLANK TEST	0.00	02:10
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:10
BLANK TEST	0.00	02:11
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:12
BLANK TEST	0.00	02:12

