

SC.UTION CHANGE




GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/03/2019

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:04
INTERNAL STANDARD	VERIFIED	01:04
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:05
BLANK TEST	0.00	01:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:06
BLANK TEST	0.00	01:07
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:08



SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 11/03/2019

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19180

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:36
INTERNAL STANDARD	VERIFIED	00:36
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:37

Blank Error

