

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 12/16/2019

OPERATOR NAME:
MARC J BARRETT
Permit #: 46725
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19250

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:14
INTERNAL STANDARD	VERIFIED	00:14
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:15
BLANK TEST	0.00	00:16
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:16
BLANK TEST	0.00	00:17
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	00:17
BLANK TEST	0.00	00:18

