

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 02/17/2020

OPERATOR NAME:
JAMES F NEAL
Permit #: 47777
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:37
INTERNAL STANDARD	VERIFIED	02:38
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:38
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:40
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:41

