

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/16/2020

OPERATOR NAME:
PATRICK J COYNE
Permit #: 46590
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:32
INTERNAL STANDARD	VERIFIED	07:32
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:33
BLANK TEST	0.00	07:34
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:34
BLANK TEST	0.00	07:35
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:35
BLANK TEST	0.00	07:36

