

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 03/23/2020

OPERATOR NAME:
MARC J BARRETT
Permit #: 46725
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 19350

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:06
INTERNAL STANDARD	VERIFIED	01:06
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:07
BLANK TEST	0.00	01:08
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:08
BLANK TEST	0.00	01:09
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	01:09
BLANK TEST	0.00	01:10

