

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/04/2020

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:41
INTERNAL STANDARD	VERIFIED	03:41
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:41
BLANK TEST	0.00	03:42
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	03:42
BLANK TEST	0.00	03:43
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.10	03:44
BLANK TEST	0.00	03:45

