

SOLUTION CHANGE

 GLENS FALLS CITY
 POLICE DEPT
 DATAMASTER dmt: 117106

Date: 05/05/2020

OPERATOR NAME:
 SHAWN A LOVELACE
 Permit #: 30657
 Operator Agency/Dept: GLENS FALLS CITY PD
 Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	14:56
INTERNAL STANDARD	VERIFIED	14:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:56
BLANK TEST	0.00	14:57
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:58
BLANK TEST	0.00	14:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	14:59
BLANK TEST	0.00	15:00

