

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 05/25/2020

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20040

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:44
INTERNAL STANDARD	VERIFIED	00:44
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:44
BLANK TEST	0.00	00:45
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:46
BLANK TEST	0.00	00:46
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:47
BLANK TEST	0.00	00:48

@ " ◆ # ▶ ◆

