

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/01/2020

OPERATOR NAME:
NICHOLAS K RYSEDORPH
Permit #: 48769
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:44
INTERNAL STANDARD	VERIFIED	02:44
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:45
BLANK TEST	0.00	02:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.10	02:46
BLANK TEST	0.00	02:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:47
BLANK TEST	0.00	02:48

