

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/15/2020

OPERATOR NAME:
NICHOLAS K RYSEDORPH
Permit #: 48769
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:56
INTERNAL STANDARD	VERIFIED	02:56
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:57
BLANK TEST	0.00	02:58
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:58
BLANK TEST	0.00	02:59
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	02:59
BLANK TEST	0.00	03:00

