

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 06/22/2020

OPERATOR NAME:
THOMAS S PRATT
Permit #: 48766
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:20
INTERNAL STANDARD	VERIFIED	00:20
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:21
BLANK TEST	0.00	00:22
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:22
BLANK TEST	0.00	00:23
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	00:23
BLANK TEST	0.00	00:24

