

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/01/2020

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:36
INTERNAL STANDARD	VERIFIED	03:36
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:36
BLANK TEST	0.00	03:37
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:38
BLANK TEST	0.00	03:39
REFERENCE STANDARD TEMP	34.03c	
REFERENCE STANDARD	0.09	03:39
BLANK TEST	0.00	03:40

