

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/13/2020

OPERATOR NAME:
DESMOND L LYONS
Permit #: 35876
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:44
INTERNAL STANDARD	VERIFIED	01:45
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:45
BLANK TEST	0.00	01:46
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:46
BLANK TEST	0.00	01:47
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:47
BLANK TEST	0.00	01:48

