

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 07/27/2020

OPERATOR NAME:
NICHOLAS L RYSEDORPH
Permit #: 48769
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20170

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:38
INTERNAL STANDARD	VERIFIED	02:38
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:39
BLANK TEST	0.00	02:39
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:40
BLANK TEST	0.00	02:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:41
BLANK TEST	0.00	02:42

