

SOLUTION CHANGE

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/01/2020

OPERATOR NAME:
GERALD C WILLETTE
Permit #: 32550
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	03:02
INTERNAL STANDARD	VERIFIED	03:02
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:02
BLANK TEST	0.00	03:03
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:04
BLANK TEST	0.00	03:04
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	03:05
BLANK TEST	0.00	03:06

