

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 08/10/2020

OPERATOR NAME:  
ROBERT M FOO  
Permit #: 46059  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:30
INTERNAL STANDARD	VERIFIED	00:30
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:30
BLANK TEST	0.00	00:31
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:31
BLANK TEST	0.00	00:32
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:32
BLANK TEST	0.00	00:33

