

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 08/24/2020

OPERATOR NAME:
THOMAS S PRATT
Permit #: 48766
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20240

Nominal = 0.10 g/210L

BLANK TEST	0.00	00:40
INTERNAL STANDARD	VERIFIED	00:40
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:41
BLANK TEST	0.00	00:41
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:42
BLANK TEST	0.00	00:43
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	00:43
BLANK TEST	0.00	00:44

