

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 10/12/2020

OPERATOR NAME:
ROBERT M FOO
Permit #: 46059
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	01:50
INTERNAL STANDARD	VERIFIED	01:51
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:51
BLANK TEST	0.00	01:52
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:52
BLANK TEST	0.00	01:53
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	01:53
BLANK TEST	0.00	01:54

