

SUPERVISOR TEST

GLENS FALLS CITY
POLICE DEPT
DATAMASTER dmt: 117106

Date: 10/26/2020

OPERATOR NAME:
SHAWN A LOVELACE
Permit #: 30657
Operator Agency/Dept: GLENS FALLS CITY PD
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	07:20
INTERNAL STANDARD	VERIFIED	07:20
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:20
BLANK TEST	0.00	07:21
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:22
BLANK TEST	0.00	07:23
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	07:23
BLANK TEST	0.00	07:24

