

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 10/05/2020

OPERATOR NAME:  
DESMOND L LYONS  
Permit #: 35876  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 20340

Nominal = 0.10 g/210L

|                         |          |       |
|-------------------------|----------|-------|
| BLANK TEST              | 0.00     | 02:53 |
| INTERNAL STANDARD       | VERIFIED | 02:54 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 02:54 |
| BLANK TEST              | 0.00     | 02:55 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 02:55 |
| BLANK TEST              | 0.00     | 02:56 |
| REFERENCE STANDARD TEMP | 34.01c   |       |
| REFERENCE STANDARD      | 0.09     | 02:57 |
| BLANK TEST              | 0.00     | 02:57 |

