

SUPERVISOR TEST

-----  
GLENS FALLS CITY  
POLICE DEPT  
DATAMASTER dmt: 117106  
-----

Date: 11/16/2020

OPERATOR NAME:  
ROBERT M FOO  
Permit #: 46059  
Operator Agency/Dept: GLENS FALLS CITY PD  
Reference Standard #: 20340

Nominal = 0.10 g/210L

BLANK TEST	0.00	02:12
INTERNAL STANDARD	VERIFIED	02:12
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:12
BLANK TEST	0.00	02:13
REFERENCE STANDARD TEMP	34.01c	
REFERENCE STANDARD	0.09	02:14
BLANK TEST	0.00	02:15
REFERENCE STANDARD TEMP	34.02c	
REFERENCE STANDARD	0.09	02:15
BLANK TEST	0.00	02:16

